



DOSIMETRY CALIBRATION LABORATORY



Nuclear Technology Service Center, Thailand Institute of Nuclear Technology (Public Organization)

9/9 Moo 7, Saimoon Sub-district, Ongkharak District, Nakorn Nayok 26120, Thailand

Tel. 02-4019889 ext. 1910, 1142, E-mail: calibration@tint.or.th, www.tint.or.th

Certificate No: GPD0054/201025

Reference No: CL0054/081025

CALIBRATION CERTIFICATE

This is to certify that the instrument described below has been calibrated by Dosimetry Calibration Laboratory, Thailand Institute of Nuclear Technology (Public Organization).

Owner:

Center for Safety,

Address:

Health and Environment of Chulalongkorn University

Chamchuri 1 Building, 1st Floor, Room 108, Phaya Thai Rd.,
Wang Mai, Pathum Wan, Bangkok 10330

Instrument:

Personal Dosimeter

Manufacturer:

RAE SYSTEMS, INC.

Model:

PRM-1200

Serial No:

0352005843

Date of receipt: 8 October 2025

Date of calibration: 20 October 2025

The calibration is traceable to the Physikalisch-Technische Bundesanstalt (PTB), the Federal Republic of Germany, through the Certificate No. 6.25-33/22K.

Approved by:

(Mr. Jeerawat Esor)



Calibrated by: Wisarut Ketaiam

Verified by: Busayakorn Na Ranong

Issued date: 22 October 2025

This certificate applies only to the identified dosimeter/contamination monitor, and shall not be reproduced except in full, and only when with written approval.



DOSIMETRY CALIBRATION LABORATORY



Nuclear Technology Service Center, Thailand Institute of Nuclear Technology (Public Organization)

9/9 Moo 7, Saimoon Sub-district, Ongkharak District, Nakorn Nayok 26120, Thailand

Tel. 02-4019889 ext. 1910, 1142, E-mail: calibration@tint.or.th, www.tint.or.th

Certificate No: GPD0054/201025

Reference No: CL0054/081025

CALIBRATION CERTIFICATE

Measurement Setup

Radiation beam: Cs-137 radioactive source
Field size: Ø 34 cm at distance 100 cm from source
Calibration method: ISO 4037:2019
Personal Dose Equivalent Rate ($\dot{H}_p(10)$): $3.363 \pm 0.216 \text{ mSv/h}$
Calibration condition: With ISO slab phantom and 3 mm PMMA plate

Standard Dosimeters/Materials

Description	Model	Serial No.	Manufacturer
Ionization Chamber	A6	XQ111652	Standard Imaging
Electrometer	Supermax	R170815	Standard Imaging

Calibration Results

Range	Standard Output (μSv)	Instrument Reading (μSv)	Calibration Factor	Uncertainty (%)
Auto	100.00	94.63	1.07	8.6
Auto	200.00	190.7		
Auto	300.00	278.6		
Auto	400.00	375.4		
Auto	500.00	466.2		

Average reading before adjustment:

The uncertainties of calibration were based on a confidence level of approximately 95% corresponding to a coverage factor of 2 ($k=2$).

Remark: -

Laboratory Environment

During calibration, the environment in calibration room was maintained within the operating specifications of the instrument and standard as following:

Relative humidity: (51.9 – 55.0) %
Ambient temperature: (21.6 – 22.1) °C
Atmospheric pressure: (1005.9 – 1006.1) hPa



46d92cf5

This certificate applies only to the identified dosimeter/contamination monitor, and shall not be reproduced except in full, and only when with written approval.