



# DOSIMETRY CALIBRATION LABORATORY



NSC-TISI-TIS 17025  
CALIBRATION 0278

Nuclear Technology Service Center, Thailand Institute of Nuclear Technology (Public Organization)

9/9 Moo 7, Saimoon Sub-district, Ongkharak District, Nakorn Nayok 26120, Thailand

Tel. 02-4019889 ext. 1910, 1142, E-mail: calibration@tint.or.th, www.tint.or.th

Certificate No: GSM0124/211025

Reference No: CL0054/081025

## CALIBRATION CERTIFICATE

This is to certify that the instrument described below has been calibrated by Dosimetry Calibration Laboratory, Thailand Institute of Nuclear Technology (Public Organization).

Owner:

Center for Safety,

Address:

Health and Environment of Chulalongkorn University

Chamchuri 1 Building, 1<sup>st</sup> Floor, Room 108, Phaya Thai Rd.,

Wang Mai, Pathum Wan, Bangkok 10330

Instrument:

Survey Meter

Manufacturer:

TINT

Model:

5702-E

Serial No:

072

Date of receipt: 8 October 2025

Date of calibration: 21 October 2025

The calibration is traceable to the Physikalisch-Technische Bundesanstalt (PTB), the Federal Republic of Germany, through the Certificate No. 6.25-33/22K.

Approved by:

(Mr. Jeerawat Esor)



Calibrated by: Wisarut Ketaiam

Verified by: Busayakorn Na Ranong

Issued date: 22 October 2025

This certificate applies only to the identified dosimeter/contamination monitor, and shall not be reproduced except in full, and only when with written approval.



# DOSIMETRY CALIBRATION LABORATORY



Nuclear Technology Service Center, Thailand Institute of Nuclear Technology (Public Organization)

9/9 Moo 7, Saimoon Sub-district, Ongkharak District, Nakorn Nayok 26120, Thailand

Tel. 02-4019889 ext. 1910, 1142, E-mail: calibration@tint.or.th, www.tint.or.th

Certificate No: GSM0124/211025

Reference No: CL0054/081025

## CALIBRATION CERTIFICATE

### Measurement Setup

Radiation beam: Cs-137 radioactive source  
 Field size: Ø 34 cm at distance 100 cm from source  
 Calibration method: ISO 4037:2019  
 Calibration condition: In air with 3 mm PMMA plate

### Standard Dosimeters/Materials

Description	Model	Serial No.	Manufacturer
Ionization Chamber	A6	XQ111652	Standard Imaging
Electrometer	Supermax	R170815	Standard Imaging

### Calibration Results

Range	Standard Output ( $\mu$ Sv/h)	Instrument Reading ( $\mu$ Sv/h)	Calibration Factor	Uncertainty (%)
x100	3363	3365	1.00	8.6
x10	326	329	0.99	8.6
x1	34.4	35	1.00	9.3

Average reading before adjustment:

x100	3363	3710	0.91
x10	326	340	0.96

The uncertainties of calibration were based on a confidence level of approximately 95% corresponding to a coverage factor of 2 ( $k=2$ ).

Remark: -

### Laboratory Environment

During calibration, the environment in calibration room was maintained within the operating specifications of the instrument and standard as following:

Relative humidity:	(48.5 – 54.3)	%
Ambient temperature:	(20.0 – 21.6)	°C
Atmospheric pressure:	(1010.2 – 1010.4)	hPa



This certificate applies only to the identified dosimeter/contamination monitor, and shall not be reproduced except in full, and only when with written approval.